



Exemption Request Form – Exemption 1b Annex IV

Date of submission: **15 January 2020**

1. Name and contact details

1) Name and contact details of applicant

Company: **COCIR** Tel.: **00327068966**
Name: **Riccardo Corridori** E-Mail: **corridori@cocir.org**
Function: **EHS Policy Senior Manager** Address: **Blvd A. Reyers 80,
1030 Bruxelles**

2) Name and contact details of responsible person for this application (if different from above):

Company: _____ Tel.: _____
Name: _____ E-Mail: _____
Function: _____ Address: _____

2. Reason for application:

Please indicate where relevant:

- Request for new exemption in:
- Request for amendment of existing exemption in
- Request for extension of existing exemption in Annex IV
- Request for deletion of existing exemption in:
- Provision of information referring to an existing specific exemption in:
 - Annex III
 - Annex IV

No. of exemption in Annex IV where applicable: **1b**

Proposed or existing wording: **Lead anodes in electrochemical oxygen sensors.**

Duration where applicable: **Until the end of 2025 for new instruments that use electrochemical oxygen sensors that contain lead. Maximum validity period for replacement oxygen sensors.**

Other: _____

3. Summary of the exemption request / revocation request

This exemption is required to allow the use of electrochemical oxygen sensors for measurement of oxygen concentrations in inhaled and exhaled air of patients who are being ventilated, and when undergoing surgery or MRI scans when under anaesthesia. Electrochemical sensors have many advantages including their very small size and no need for a power supply which provide them with unique functionality critical to patient care.

Alternative types of oxygen sensor have been assessed, but all alternative types are unsuitable for the aforementioned applications. Lead-free electrochemical sensors have recently become available and have been evaluated. Tests have shown that these are not drop-in replacements and cannot be used with the existing oxygen analyser instruments currently in use in EU hospitals and clinics. Analyser instruments that are connected to the sensors and indicate the oxygen concentration are being redesigned to use new lead-free sensors although these cannot be sold in the EU until redesign, testing a qualification is complete and Medical Device Regulation approval is granted which is not expected before 2025. This exemption will be required after 2025 to allow the currently used lead-based sensors to be used as replacements with the current designs of analyser instruments that are in use in EU hospitals and clinics.

4. Technical description of the exemption request / revocation request

(A) Description of the concerned application:

1. To which EEE is the exemption request/information relevant?

Name of applications or products: Respiratory care products that measure oxygen concentration of inhaled and exhaled air for patients, for example in anaesthesia workstations and in ventilators.

a. List of relevant categories: (mark more than one where applicable)

- | | |
|----------------------------|---------------------------------------|
| <input type="checkbox"/> 1 | <input type="checkbox"/> 7 |
| <input type="checkbox"/> 2 | <input checked="" type="checkbox"/> 8 |
| <input type="checkbox"/> 3 | <input type="checkbox"/> 9 |
| <input type="checkbox"/> 4 | <input type="checkbox"/> 10 |
| <input type="checkbox"/> 5 | <input type="checkbox"/> 11 |
| <input type="checkbox"/> 6 | |

b. Please specify if application is in use in other categories to which the exemption request does not refer: We believe that these sensors are also used in category 9 applications, but these are not included in this renewal request

c. Please specify for equipment of category 8 and 9:

The requested exemption will be applied in

monitoring and control instruments in industry

in-vitro diagnostics

other medical devices or other monitoring and control instruments than those in industry

2. Which of the six substances is in use in the application/product?

(Indicate more than one where applicable)

Pb

Cd

Hg

Cr-VI

PBB

PBDE

3. Function of the substance: Electrochemical oxygen concentration measurement

4. Content of substance in homogeneous material (%weight): >99% lead

5. Amount of substance entering the EU market annually through application for which the exemption is requested: ca 8 to 13.5 kg per year

Please supply information and calculations to support stated figure.

Each cell typically contains 8 to 9 grams of lead

Estimated sales in the EU for medical applications is expected to be about 1000 to 1500 sensors per year

6. Name of material/component: Lead metal

7. Environmental Assessment: _____

LCA: Yes

No

(B) In which material and/or component is the RoHS-regulated substance used, for which you request the exemption or its revocation? What is the function of this material or component?

Oxygen in patients' breath is measured by small electrochemical sensors that contain lead anodes. These are used very close to the patient's mouth so must be small and lightweight. These are connected electrically to much larger analyser instruments that calculate and display the oxygen concentration. The sensors are connected to the instrument using electrical cables.


The concentration of oxygen in air supplied to patients and in their exhaled breath needs to be monitored to ensure that they are not being harmed by a poor air supply and monitor their medical condition. Oxygen concentrations need to be accurately monitored if a patient is very ill (e.g. needing to be ventilated), during

surgery, incubators for infant and premature babies or if patient are anaesthetised during MRI (magnetic resonance imaging) examination. Sensors used during MRI scans must not contain magnetic materials which occur in most types of electronic components (these usually contain nickel as a barrier coating on terminations).

Lead (Pb) is used as the anode in the oxygen (O₂) sensor. During the operation of the sensor, oxygen from the gas being analysed permeates through the membrane of the sensor and is electrochemically reduced at the cathode of the sensor. The Pb anode of the sensor is oxidised to PbO / PbO₂.

Electrochemical oxygen sensors consist of two electrodes; a high surface area lead anode and an inert cathode which are immersed in an alkali electrolyte. The anode and cathode reactions are:

Cathode:



Anode:



The electrochemical reaction generates an electrical current which flows through the cell (with an additional load resistor) which is proportional to the partial pressure of O₂ in the analysed gas. The voltage across the load resistor is proportional to the current and is measured to calculate the concentration of O₂.

The output current of the electrochemical sensor is therefore dependent on the oxygen concentration in the air that enters the sensor. The rate of access of air-oxygen into the sensor is controlled by the design of the sensor, in particular by using diffusion barriers which can be narrow capillaries or a porous layer. The sensor is designed so that the current output is proportional to the oxygen concentration. The diffusion barrier is used to limit the amount of air that passes through the sensor so that only a small amount of oxygen reaches the lead anode so that a proportionally small amount of lead is consumed, ensuring the maximum lifetime of the product.

With the increasing conversion of lead to lead oxide, the voltage over the measurement resistor slowly decreases during use. To ensure the correct calculation of the O₂ concentration, the instrument and sensor is calibrated with air every 24 hours.

When the voltage with air drops below 8.9 mV (under normal conditions) the medical device detects that the sensor is consumed and informs the user to replace the sensor with a new one. Under normal conditions the sensor can be used for 1 to 2 years.

Ambient temperature also affects the current output of the sensor and so compensation circuits need to be included either within the sensor or in the current

measurement instrument. This is straightforward with lead-based sensors, but is also a critical variable with other types of sensor.

Accurate O₂-measurement in breathing gases (inhalation / exhalation) is a prerequisite regarding patient safety for the correct control of anaesthesia procedures. Beside the requirement regarding measurement accuracy, rapid response behaviour is also essential and this analysis method of O₂-measurement ensures immediate reaction strategies by the anaesthetist during medical surgery.

One of the main advantages of lead in electrochemical oxygen sensors is that it is not affected by the majority of other gases used in medical applications, such as with patient ventilators or during anaesthesia. The few exceptions to this are acid gases and CO₂ will react with the electrolyte (potassium hydroxide or potassium acetate are usually used) and users are warned that this should be minimised, but anaesthesia gases have only a minimal effect on this type of sensor.

(C) What are the particular characteristics and functions of the RoHS-regulated substance that require its use in this material or component?

For monitoring patients, the following characteristics are essential:

- Fast response. For example, <15 seconds to respond to a change from air to 100% oxygen
- High sensitivity to small oxygen concentration changes
- High accuracy
- Long lifetime (at least 1 year, ideally 2 years)
- Measure oxygen in the concentration range of 0% to 100%
- Must be usable and accurate in the presence of commonly used anaesthesia gases (such as Halothane, Isoflurane, Sevoflurane and Desflurane), N₂O and CO₂.
- Usable at 0% – 99% non-condensing humidity
- Accuracy should not be affected by typical CO₂ concentrations in exhaled breath
- Sensors must not contain magnetic materials when they are used with MRI scanners

5. Information on Possible preparation for reuse or recycling of waste from EEE and on provisions for appropriate treatment of waste

1) Please indicate if a closed loop system exist for EEE waste of application exists and provide information of its characteristics (method of collection to ensure closed loop, method of treatment, etc.)

No, used sensors are recycled for materials recovery.

2) Please indicate where relevant:

- Article is collected and sent without dismantling for recycling
- Article is collected and completely refurbished for reuse
- Article is collected and dismantled:
 - The following parts are refurbished for use as spare parts: _____
 - The following parts are subsequently recycled: _____
- Article cannot be recycled and is therefore:
 - Sent for energy return
 - Landfilled

3) Please provide information concerning the amount (weight) of RoHS substance present in EEE waste accumulates per annum:

- In articles which are refurbished _____
- In articles which are recycled 8 to 13.5 kg per year
- In articles which are sent for energy return _____
- In articles which are landfilled _____

6. Analysis of possible alternative substances

(A) Please provide information if possible alternative applications or alternatives for use of RoHS substances in application exist. Please elaborate analysis on a life-cycle basis, including where available information about independent research, peer-review studies development activities undertaken

Since categories 8 and 9 were included in scope of RoHS, research into substitutes for lead in electrochemical sensors has been carried out. There are also different types of oxygen sensor on the market but these operate in completely different ways to electrochemical sensors and so have very different characteristics. Both alternative metals to lead and alternative types of sensor are described here.

Lead substitutes in electrochemical sensors

Research has been carried out and published with alternative anode metals and with various acid and alkali electrolytes. Metals that have been investigated include: antimony, bismuth, copper, tin and its alloys, zinc and aluminium.

Research has shown¹ that the more reactive metals such as tin, zinc and aluminium are unsuitable as they are thermodynamically unstable in suitable electrolytes. Electrolytes such as potassium hydroxide (used with lead) and other alkali solutions, acidic solutions such as phosphoric acid and caesium carbonate solution (mildly alkali) have been investigated.

When a reactive metal anode is combined with an inert cathode, the two different materials generate a galvanic couple which creates a small voltage with the electrode potential of the anode such that they self-corrode, generating a current and generate hydrogen. This generated current between anode and cathode gives a false and incorrect oxygen concentration. Despite this, commercially available oxygen sensors that use tin anodes in a caesium carbonate electrolyte² are being sold in the EU³. The applicants of this renewal request have however evaluated these sensors for their specific applications and found that they do not meet their specifications when used with existing designs of analyser instrument.

Research with less reactive metals such as copper, bismuth and antimony has also been reported, but no commercial products have been developed. The reason why these metals are not used in commercial sensors could be that they can form thin oxide coatings (e.g. during storage before use) which may act as a barrier to further oxidation and so hinder or prevent further electrochemical reaction.

Lead may be the optimal anode choice in electrochemical sensors because it does not self-corrode in the absence of oxygen (such as aluminium and zinc) but it responds rapidly when in contact with oxygen, unlike copper, which reacts and then rapidly passivates so stops working. Noble metals such as gold and silver do not respond at all as they do not react with oxygen from air.

When used for monitoring breathing when a patient is anaesthetised, it is important that the accuracy of the sensor is not affected by anaesthetic gases. Many years of use in hospitals has shown that these gases have no effect on lead in electrochemical sensors, but extensive testing is needed with any potential substitutes to determine whether it will be accurate and reliable before

¹ Lead-Free Galvanic Oxygen Sensors. A Conceptual Approach, Cornel Cobianu, et. al (Honeywell). CAS (International Semiconductor Conference) 2012, Abstract from <https://www.semanticscholar.org/paper/Lead-free-galvanic-oxygen-sensors-%E2%80%94-A-conceptual-Cobianu-Serban/e4cfb461b42eba465ee2410d5637bf7453079bf6>

² <http://www.it-wismar.de/download.php?id=22>

³ Marketed by ITG.

approval can be obtained under the Medical Devices Regulation. Anaesthesia gases are water soluble to a small extent, so small amounts will dissolve in the sensor's electrolyte and could affect its function.

Additionally when some anaesthetic gases, such as halothane, are exposed to sunlight or UV light they decompose to give halides which may affect the way that the anode dissolves. Therefore it cannot be assumed that alternative metals to lead will respond in a similar way when used in oxygen sensors for medical applications and so extensive testing is essential.

Tests with a type of commercial lead-free electrochemical sensor have been carried out by one manufacturer. Concerning the output signal itself, results indicate that sensors with alternative anode materials should eventually be suitable as a replacement for the lead anode sensors in redesigned instruments. However, test results showed that these sensors have a completely different behaviour in the way that they decrease voltage during operation. The voltage decreases very slowly and does not drop below 8.9 mV even if the anode is completely consumed.


In existing designs of oxygen analyser, a consumed sensor with this behaviour would lead to wrong O₂-values with no possibility of detection by the device or the user that the output data was not correct.

Due to the internal design of the cell (which is required by the anode material) it is not possible to adjust the way that the output voltage changes as the anode is consumed by electrical means to reproduce the behaviour of a cell with a lead anode. Therefore lead-free sensors cannot be used as drop-in replacements for the lead anode sensors used in currently available commercial oxygen analysers. The timescale required for instrument redesign is explained in section 7.

Other types of oxygen sensor

Many methods of measuring oxygen concentration have been developed but each has different characteristics. These methods are described here:

3-electrode electrochemical sensors. One design that has been developed is three electrode cells containing inert lead-free anodes and cathodes with a reference electrode. These have the advantage that the anode is not consumed as the anode reaction is:



These however have the disadvantage that oxygen is generated at the anode and must not reach the cathode as this would give a false high reading. The

accuracy of these sensors is also affected by ambient temperature, humidity, pressure and the presence of other gases such as anaesthetics.

One advantage of the lead anode sensors is that they carry their “own” voltage power supply in the anode material (which is the reason for the sensors not working when the anode material is consumed). Electrochemical 3-electrode O₂ - sensors require a bipotentiostat which needs an additional power supply. This is not available at the installation location of the sensor due to a lack of space in the part of the medical device where the sensor is located as this needs to be fairly close to the patient’s mouth. Also, the capacity of a small enough battery that would fit into the available space would not allow the operation over 1-2 years.

Experiments with this sensor also showed that electromagnetic requirements cannot be fulfilled without additional means, which are problematic at the installation place of the sensor (e.g. metal housing and control electronics). These also cannot be used with MRI as the control circuits include magnetic materials.

Solid metal oxide semiconductor sensors

These are known as lambda probes and are often used in vehicles to measure the differences in oxygen concentrations in supplied air and in exhaust gases by measurement of the current between electrodes. They do not measure the actual oxygen concentration (only a difference) and these usually need to be heated to 300°C and so are unsuitable in medical applications as they pose a safety risk due to the high temperature.

Mass spectroscopy

Mass spectrometers can analyse all substances in patients’ breath but the response time is too long taking many minutes to obtain a single concentration⁴.

Paramagnetic oxygen gas analysers

These rely on oxygen gas being strongly paramagnetic and this method can give very fast response times. However, they have several disadvantages. The response is affected by the concentrations of other paramagnetic gases which includes CO₂ (in exhaled breath), N₂O (used for anaesthesia) and most other anaesthesia gases. It is also affected by water vapour concentrations (water is diamagnetic so opposes the paramagnetic effect⁵). Because of these limitations, paramagnetic oxygen sensors are unsuitable in anaesthesia applications and can be difficult to use with ventilators.

Paramagnetic sensors have already replaced lead anode based O₂ measurement in diverting (side-stream) respiratory gas monitors. However, due

⁴ <https://academic.oup.com/bjaed/article/9/1/19/465989>

⁵ https://www.sablesys.com/wp-content/uploads/Sable-Systems-International_White_Paper_O2-Analyzer.pdf

to requirements for the flow of the measurement gas to be very continuous, without pressure and flow peaks, it is not possible to use this type of sensor for the inspiratory O₂ measurement for anaesthesia applications or in main-stream ventilators. This is because the flow rates in the inspiratory part of the devices has a big range. Another disadvantage of this type of sensor is that they need an additional power supply, which cannot be made available near the sensor due to a lack of space. The current peaks and power needed by this principle would require quite a large battery, which is mechanically and dimensionally not possible at the installation location of the sensor.

Optical sensors

There are many types of optical oxygen concentration analysis methods, but most are either unsuitable or are designed for analysis of dissolved oxygen only (e.g. in blood)⁶. Infrared absorption spectroscopy analysis can be used to analyse many substances in breath including anaesthesia concentrations, but this technique does not respond to oxygen gas and so is unsuitable.

One technique uses substances that change colour when exposed to oxygen and the colour is measured electronically. Response times can be slow and colour measurement is not very accurate and so this method cannot be used for ventilators or anaesthesia oxygen monitors where fast response times and high accuracy are essential.

Another type of optical sensor uses an oxygen permeable polymer containing a luminescent compound. The luminescent compounds emit light when excited by exposure to light of a shorter wavelength (e.g. UV), but this is quenched when exposed to oxygen and the change in luminescence intensity is used as a measure of oxygen concentration.

Many luminescent oxygen sensors have been developed and a few types have been commercialised. A recent review⁷ reports that many luminescent dyes can be used in aqueous solutions, but some are effective only within specific pH ranges, so are used only to measure dissolved oxygen (not the gas). Many types of dye are not stable as they photo-bleach and become unresponsive. Response time of some types is a disadvantage; one commercial product is claimed to have a response time of less than 30 seconds⁸, however electrochemical sensors respond typically in less than 10 seconds and some are <5 seconds. A fast response time is essential for monitoring any changes in breathing of ventilated patients and patients under anaesthesia. A few types of luminescent oxygen

⁶ <https://www.optimedical.com/>

⁷ Optical methods for sensing and imaging oxygen: materials, spectroscopies and applications, Xu-dong Wang and Otto S. Wolfbeis, Chem. Soc. Rev., 2014, 43, 3666.

⁸ Luminox O2 sensors available from SST (technical datasheet)

sensors have much faster response times but only function at very low oxygen concentrations that are not applicable to medical applications.

Luminescent oxygen sensors respond by quenching by oxygen molecules, however many other substances including anaesthesia gases, water vapour, nitrogen oxides and many other contaminants also causes quenching and so these sensors are not sufficiently selective for oxygen to be useful in medical ventilation and anaesthesia applications.

(B) Please provide information and data to establish reliability of possible substitutes of application and of RoHS materials in application

Reliability is not believed to be an issue and substitutes are technically unsuitable as explained in section 6 (A).

7. Proposed actions to develop possible substitutes

(A) Please provide information if actions have been taken to develop further possible alternatives for the application or alternatives for RoHS substances in the application.

Research has been carried and published on a wide variety of new lead-free oxygen sensors. Medical device manufacturers evaluate any commercial products that meet their specifications and appear to be suitable. As stated above, paramagnetic sensors are now used in side-stream applications, but cannot be used where rapid and accurate measurements are required such as in anaesthesia applications.

New lead-free electrochemical oxygen sensors are being evaluated and tests have shown that these respond very differently and so are not drop-in replacements. As a result, analyser instruments that are used with the sensors will have to be completely redesigned. As these are safety critical medical devices, any new designs will be required to undergo reliability and accuracy testing followed by clinical trials before the technical file can be submitted to an EU Notified Body for Medical Devices Regulation approval and before these can be sold in the EU.

(B) Please elaborate what stages are necessary for establishment of possible substitute and respective timeframe needed for completion of such stages.

Medical device manufacturers are reliant on manufacturers of sensors to develop lead-free substitutes that are accurate and have fast response when used in

ventilators and anaesthesia workstations. For MRI applications, they must also not contain magnetic metals such as nickel and iron.

Medical device manufacturers are evaluating new sensors that appear to meet the specifications, but redesign of analysers is also required followed by further extensive testing for accuracy and reliability is needed before they can be evaluated using clinical trials. Once these trials are successfully completed, Notified Body approval in the EU and the equivalent approvals globally can be requested. Typical timescales would be once a suitable sensor becomes available for testing:

- Testing of sensor 1 year (currently underway)
- Redesign of analysers 1 year
- Testing in ventilators and in anaesthesia workstations 6 months
- Clinical trials ca. 1 year
- Global approvals Up to 2 years
- Total elapsed time Up to about 5.5 years

In theory, if no difficulties are encountered, analysers suitable for lead-free sensors should be available by the end of 2025, however, EU hospitals and clinics will not be able to use replacement lead-free sensors in current designs of analysers as these function correctly only with lead anode sensors and so the exemption will continue to be required for replacement sensors after this date⁹.

⁹ Unless the lead restriction does not apply due to Article 4.4.

8. Justification according to Article 5(1)(a):

(A) Links to REACH: (substance + substitute)

1) Do any of the following provisions apply to the application described under (A) and (C)?

Authorisation

SVHC

Candidate list

Proposal inclusion Annex XIV

Annex XIV

Restriction

Annex XVII

Registry of intentions

Registration lead has been registered – see <https://ila-reach.org/our-substances/lead-metal/> and <https://echa.europa.eu/registration-dossier/-/registered-dossier/16063>

2) Provide REACH-relevant information received through the supply chain.

Name of document: _____

(B) Elimination/substitution:

1. Can the substance named under 4.(A)1 be eliminated?

Yes. Consequences? _____

No. Justification: None of the potential substitutes achieve all of the essential performance criteria

2. Can the substance named under 4.(A)1 be substituted?

Yes.

Design changes:

Other materials:

Other substance:

No.

Justification: None of the potential substitutes achieve all of the essential performance criteria

3. Give details on the reliability of substitutes (technical data + information): The substitutes are technically unsuitable therefore reliability is unable to be investigated.

4. Describe environmental assessment of substance from 4.(A)1 and possible substitutes with regard to; Not applicable to this exemption renewal request as substitute is technically unsuitable

- 1) Environmental impacts: _____
- 2) Health impacts: _____
- 3) Consumer safety impacts: _____

⇒ Do impacts of substitution outweigh benefits thereof? Not applicable to this exemption renewal request

Please provide third-party verified assessment on this: _____

(C) Availability of substitutes:

- a) Describe supply sources for substitutes: Lead-free oxygen sensors have recently become available but are not drop-in replacements, as explained in section 6
- b) Have you encountered problems with the availability? Describe: No as there are no drop-in alternatives available
- c) Do you consider the price of the substitute to be a problem for the availability?
 Yes No
- d) What conditions need to be fulfilled to ensure the availability? See sections 6 and 7

(D) Socio-economic impact of substitution:

⇒ What kind of economic effects do you consider related to substitution?

- Increase in direct production costs
- Increase in fixed costs
- Increase in overhead

Possible social impacts within the EU: EU hospital patients who undergo surgery under anaesthesia, patients being ventilated and patients under anaesthesia having MRI scans would be at risk of harm or death if the electrochemical oxygen sensors that contain lead anodes could no longer be supplied to EU hospitals. There are published statistics on the numbers of surgical procedures carried out annually¹⁰ and most of these rely on anaesthetics. For example, some operations such as cataracts and caesarean sections usually do not use gaseous anaesthetics, but many types of operation usually require use of anaesthesia, such as appendectomies and hip replacements. One study found that 3.2 million people underwent surgery under anaesthesia in the UK in 2013¹¹, so the number for the EU is likely to be >10 million people per year and so this number of EU citizens could be negatively affected if this

¹⁰ https://ec.europa.eu/eurostat/statistics-explained/index.php/Surgical_operations_and_procedures_statistics

¹¹ <https://www.roca.ac.uk/news-and-bulletin/roca-news-and-statements/nationwide-survey-of-all-uk-hospitals-shows-anaesthetic>

exemption is not renewed. Oxygen measurement using less suitable methods during surgery would increase the risk of harm, potentially including deaths.

Possible social impacts external to the EU

Other: _____

⇒ Provide sufficient evidence (third-party verified) to support your statement: _____

9. Other relevant information

Please provide additional relevant information to further establish the necessity of your request:

10. Information that should be regarded as proprietary

Please state clearly whether any of the above information should be regarded to as proprietary information. If so, please provide verifiable justification:
